



Technical Data

Fluid Tetrathionate Medium w/o Iodine and BG (Tetrathionate Broth Base w/o Iodine and BG) M032

Intended Use:

Fluid Tetrathionate Medium with added iodine and brilliant green is recommended for the selective enrichment method for isolating *Salmonellae* from faeces, urine, food and other material of sanitary importance.

Composition**

Ingredients	Gms / Litre
Tryptone	2.500
Peptone	2.500
Bile salts	1.000
Calcium carbonate	10.000
Sodium thiosulphate	30.000

**Formula adjusted, standardized to suit performance parameters

Directions

Suspend 46.0 grams in 1000 ml distilled water and heat just to boiling. DO NOT AUTOCLAVE. Cool below 45°C and add 20 ml iodine solution (iodine - 6 grams and potassium iodide - 5 grams in 20 ml distilled water) and 10 ml of 0.1% brilliant green solution. Mix well and dispense in 10 ml quantities. This complete medium should be used on the day of preparation otherwise sterilized broth base may be stored for some time. Do not heat after the addition of iodine solution. Use the medium immediately after addition of iodine.

Note: Due to the presence of calcium carbonate, the prepared medium forms opalescent solution with white precipitate.

Principle And Interpretation

Salmonella are ubiquitous in the environment. These organisms are usually present in small numbers compared to coliforms; therefore it is necessary to examine a relatively large sample to isolate the organisms (1). *Salmonella* present in food samples may be sublethally damaged during various stages of food processing where they may be exposed to low temperatures, heat drying, radiations, various chemicals (2). These damaged cells are able to cause spoilage, and if ingested cause diseases under favourable conditions. Therefore it is important to resuscitate these damaged bacteria before enumeration. Fluid Tetrathionate Medium (with added iodine and brilliant green) is recommended for the selective enrichment of *Salmonella* including *Salmonella* Typhi from faeces, urine, food and other material of sanitary importance. The medium, originally formulated by Mueller (3) is recommended by APHA (4-6) for enrichment of *Salmonella*.

Due to the addition of iodine and potassium iodide, tetrathionate is formed in the medium. Organisms possessing the enzyme tetrathionate reductase grow in this medium.

Casein enzymic hydrolysate and peptic digest of animal tissue are the sources of carbon, nitrogen, vitamins and minerals. Bile salts inhibit accompanying gram-positive microorganisms. The selectivity depends on the ability of thiosulphate and tetrathionate in combination to suppress commensal coliform organism (7, 8). Calcium carbonate neutralizes the acidic tetrathionate decomposition products. Brilliant green also helps to select *Salmonella* by inhibiting the accompanying bacteria. For further confirmation, streak the enriched cultures after incubation, on plates of Brilliant Green Agar (M016), MacConkey Agar (M081) and Bismuth Sulphite Agar (M027).

Aseptically inoculate test specimen into Fluid Tetrathionate medium (with added iodine and brilliant green) and incubate at 35-37°C for 18-24 hours. Following the incubation, isolate onto selective media plates. Refer standard procedures for enrichment and isolation (4-6)

Type of specimen

Please refer disclaimer Overleaf.

NA

Specimen Collection and Handling

For clinical samples follow appropriate techniques for handling specimens as per established guidelines (4,5). For food and dairy samples, follow appropriate techniques for sample collection and processing as per guidelines (1,2,8). For water samples, follow appropriate techniques for sample collection, processing as per guidelines and local standards.(3) After use, contaminated materials must be sterilized by autoclaving before discarding.

Warning and Precautions

In Vitro diagnostic use only. Read the label before opening the container. Wear protective gloves/protective clothing/eye protection/face protection. Follow good microbiological lab practices while handling specimens and culture. Standard precautions as per established guidelines should be followed while handling clinical specimens. Safety guidelines may be referred in individual safety data sheets.

Limitations

NA

Performance and Evaluation

Performance of the medium is expected when used as per the direction on the label within the expiry period when stored at recommended temperature.

Quality Control

Appearance

White to cream homogeneous free flowing powder

Colour and Clarity of prepared medium

Complete medium with added brilliant green and iodine solution - Light green coloured, opalescent solution with heavy white precipitate, which on standing the precipitate settles down.

Cultural Response

Cultural characteristics observed with added brilliant green and iodine solution when sub cultured on MacConkey Agar (M081) after enrichment in Tetrathionate medium, after an incubation at 35-37°C for 18-24 hours.

Cultural Response

Organism	Inoculum (CFU)	Recovery	Colour of colony
Cultural Response			
<i>Escherichia coli</i> ATCC 25922 (00013*)	50-100	little or no increase in number	pink-red with bile precipitate
<i>Salmonella Choleraesuis</i> ATCC 12011	50-100	good-excellent	colourless
<i>Salmonella Typhi</i> ATCC 6539	50-100	good-excellent	colourless
<i>Salmonella Typhimurium</i> ATCC 14028 (00031*)	50-100	good-excellent	colourless
<i>Escherichia coli</i> NCTC 9002	50-100	little or no increase in numbers	pink-red with bile precipitate
<i>Escherichia coli</i> ATCC 8739 (00012*)	50-100	little or no increase in numbers	pink-red with bile precipitate

Key : *Corresponding WDCM numbers.

Storage and Shelf Life

Store below 30°C in a tightly closed container and the prepared medium at 2 - 8°C. Use before expiry date on the label. On opening, product should be properly stored dry, after tightly capping the bottle in order to prevent lump formation due to the hygroscopic nature of the product. Improper storage of the product may lead to lump formation. Store in dry ventilated area

protected from extremes of temperature and sources of ignition Seal the container tightly after use. Use before expiry date on the label. Product performance is best if used within stated expiry period.

Disposal

User must ensure safe disposal by autoclaving and/or incineration of used or unusable preparations of this product. Follow established laboratory procedures in disposing of infectious materials and material that comes into contact with clinical sample must be decontaminated and disposed of in accordance with current laboratory techniques (6,7).

Reference

1. Cherry et al, 1972, Appl. Microbiol., 24:334
2. Hartman and Minich, 1981, J. Food and Prot., 44:385
3. Mueller, 1923, Compt. Rend. Sco. Biol., 89:434.
4. Downes F. P. and Ito K., (Eds.), 2001, Compendium of Methods For The Microbiological Examination of Foods, 4th Ed., APHA, Washington, D.C.
5. Eaton A. D., Clesceri L. S. and Greenberg A. W., (Eds.), 2005, Standard Methods for the Examination of Water and Wastewater, 21st Ed., APHA, Washington, D.C.
6. FDA Bacteriological Analytical Manual, 2005, 18th Ed., AOAC, Washington, DC.
7. Pollock M. R. and Knor R., 1943, Biochem J., 37:476.
8. MacFaddin J. F., 1985, Media for Isolation-Cultivation-Identification-Maintenance of Medical Bacteria, Vol. 1, Williams and Wilkins, Baltimore.


Revision : 02 / 2016



In vitro diagnostic medical device



CE Marking



Storage temperature



Do not use if package is damaged



HiMedia Laboratories Pvt. Limited,
23 Vadhani Industrial Estate,
LBS Marg, Mumbai-86, MS, India



CE Partner 4U ,Esdoornlaan 13, 3951
DB Maarn The Netherlands,
www.cepartner4u.eu

Disclaimer :

User must ensure suitability of the product(s) in their application prior to use. Products conform solely to the information contained in this and other related HiMedia™ publications. The information contained in this publication is based on our research and development work and is to the best of our knowledge true and accurate. HiMedia™ Laboratories Pvt Ltd reserves the right to make changes to specifications and information related to the products at any time. Products are not intended for human or animal or therapeutic use but for laboratory, diagnostic, research or further manufacturing use only, unless otherwise specified. Statements contained herein should not be considered as a warranty of any kind, expressed or implied, and no liability is accepted for infringement of any patents.