



## Performance and Evaluation

Performance of the product is expected when used as per the direction on the label within the expiry period when stored at recommended temperature.

## Quality Control

### Microscopic Examination

Negative staining was carried out and observed under oil immersion lens.

### Results

Capsule: Clear halos against dark background

## Storage and Shelf Life

Store between 10- 30°C in tightly closed container and away from bright light. Use before expiry date on label. On opening, product should be properly stored in dry ventilated area protected from extremes of temperature and sources of ignition. Seal the container tightly after use.

## Disposal

User must ensure safe disposal by autoclaving and/or incineration of used or unusable preparations of this product. Follow established laboratory procedures in disposing of infectious materials and material that comes into contact with clinical sample must be decontaminated and disposed of in accordance with current laboratory techniques (1, 2).

## Reference

1. Collee J. G./Fraser A.G., Marmion B.P./Simmons A., 1996, 14th ed., MACKIE McCARTNEY, PRACTICAL MEDICAL MICROBIOLOGY: 45(800-802)
2. George Clark et al, 1981, 4th ed., Staining procedures: 17(380-382)
3. Godkar B. P., 1996, Textbook of medical laboratory technology: 23(316)

Revision : 01/ 2019



In vitro diagnostic medical device



CE Marking



Storage temperature



Do not use if package is damaged



HiMedia Laboratories Pvt. Limited,  
B /4-6 , MIDC, Dindori, Nashik MH  
[www.himedialabs.com](http://www.himedialabs.com)



CE Partner 4U ,Esdoornlaan 13, 3951  
DB Maarn The Netherlands,  
[www.cepartner4u.eu](http://www.cepartner4u.eu)

### Disclaimer :

User must ensure suitability of the product(s) in their application prior to use. Products conform solely to the information contained in this and other related HiMedia™ publications. The information contained in this publication is based on our research and development work and is to the best of our knowledge true and accurate. HiMedia™ Laboratories Pvt Ltd reserves the right to make changes to specifications and information related to the products at any time. Products are not intended for human or animal or therapeutic use but for laboratory, diagnostic, research or further manufacturing use only, unless otherwise specified. Statements contained herein should not be considered as a warranty of any kind, expressed or implied, and no liability is accepted for infringement of any patents.