



Diluting Fluid A

LQ121C

Intended use

Recommended as Diluent in testing of pharmaceuticals in accordance with USP .

Composition**

Ingredients	Gms / Litre
Peptone	1.000
Final pH (at 25°C)	7.1±0.2

**Formula adjusted, standardized to suit performance parameters

Directions

Diluting fluid is used as the diluting or rinsing solution for membrane filter techniques in pharmaceutical products. Measured portions of Diluting fluid A should be used to rinse the membrane after filtration. Inoculate this rinse with 50-100 cfu of test organisms. Simultaneously run a positive control of the same medium. Incubate both the set of medium at the specified time and temperature. Compare the growth obtained for the rinse with that obtained in the positive control after incubation.

Principle And Interpretation

Diluting Fluid A is recommended as rinsing fluid for membrane filter method used in validation tests for bacteriostasis and fungistasis activity of pharmaceutical articles before carrying out sterility test procedures as per USP (3). After filtering the specified quantity of the test specimen, the membrane is rinsed with measured portions of rinsing or diluting fluid. This rinse is inoculated with known number of test bacteria and fungi as specified in pharmacopoeia. The resultant growth is compared with positive control to determine presence of fungistasis or bacteriostasis activity in test specimen.

Type of specimen

Pharmaceutical samples

Specimen Collection and Handling

For pharmaceutical samples follow appropriate techniques for handling specimens as per established guidelines (3).

After use, contaminated materials must be sterilized by autoclaving before discarding.

Warning and Precautions

Read the label before opening the container. Wear protective gloves/protective clothing/eye protection/face protection. Follow good microbiological lab practices while handling specimens and culture. Standard precautions as per established guidelines should be followed while handling specimens. Safety guidelines may be referred in individual safety data sheets

Limitations

1. Further biochemical and serological testing must be carried out for complete identification.

Performance and Evaluation

Performance of the medium is expected when used as per the direction on the label within the expiry period when stored at recommended temperature.

Quality Control

Appearance

Sterile clear Diluting Fluid A in glass bottle

Colour

Light yellow coloured medium

Quantity of medium

100 ml of medium in glass bottle

pH

6.90-7.30

Sterility test

Passes release criteria.

Growth Promotion Test

In accordance with the harmonized method of USP.

Cultural Response

Cultural characteristics observed after an incubation at 35-37°C for 24-48 hours

Organism	Inoculum (CFU)	Growth
<i>Candida albicans</i> ATCC 10231 (00054*)	50-100	good
<i>Escherichia coli</i> ATCC 25922 (00013*)	50-100	good
<i>Staphylococcus aureus</i> subsp. <i>aureus</i> ATCC 25923 (00034*)	50-100	good
<i>Staphylococcus aureus</i> subsp. <i>aureus</i> ATCC 6538 (00032*)	50-100	good

Key: (*) Corresponding WDCM Numbers

Storage and Shelf Life

Store between 15-25°C. Use before expiry date on the label.

Disposal

User must ensure safe disposal by autoclaving and/or incineration of used or unusable preparations of this product. Follow established laboratory procedures in disposing of infectious materials and material that comes into contact with clinical sample must be decontaminated and disposed of in accordance with current laboratory techniques (1,2).

Reference

1. Isenberg, H.D. Clinical Microbiology Procedures Handbook. 2nd Edition.
2. Jorgensen, J.H., Pfaller, M.A., Carroll, K.C., Funke, G., Landry, M.L., Richter, S.S and Warnock, D.W. (2015) Manual of Clinical Microbiology, 11th Edition. Vol. 1.
3. The United States Pharmacopoeia, 2019, US Pharmacopoeial convention Inc., Rockville, MD.

Revision : 00 / 2019

Disclaimer :

User must ensure suitability of the product(s) in their application prior to use. Products conform solely to the information contained in this and other related HiMedia™ publications. The information contained in this publication is based on our research and development work and is to the best of our knowledge true and accurate. HiMedia™ Laboratories Pvt Ltd reserves the right to make changes to specifications and information related to the products at any time. Products are not intended for human or animal or therapeutic use but for laboratory, diagnostic, research or further manufacturing use only, unless otherwise specified. Statements contained herein should not be considered as a warranty of any kind, expressed or implied, and no liability is accepted for infringement of any patents.