



MUG EC Broth

M1042

Intended Use:

Recommended for detection of *Escherichia coli* in water and food samples by a fluorogenic method.

Composition**

Ingredients	Gms / Litre
Tryptone	20.000
Lactose	5.000
Bile salts mixture	1.500
Dipotassium hydrogen phosphate	4.000
Potassium dihydrogen phosphate	1.500
Sodium chloride	5.000
4-Methylumbelliferyl β-D-Glucuronide (MUG)	0.050
Final pH (at 25°C)	6.9±0.2

**Formula adjusted, standardized to suit performance parameters

Directions

Suspend 37.05 grams in 1000 ml purified / distilled water. Heat, if necessary to dissolve the medium completely. Dispense in tubes containing inverted Durham's tubes. Sterilize by autoclaving at 15 lbs pressure (121°C) for 12-15 minutes.

Principle And Interpretation

Escherichia coli is a member of faecal coliform group of bacteria. It is a member of the indigenous faecal flora of warm-blooded animals. *E.coli* is considered a specific indicator of faecal contamination and the possible presence of enteric pathogens. EC Broth was devised by Hajna and Perry (1) and further modified by addition of the fluorogenic compound MUG. MUG EC Broth is also recommended by APHA for the analysis of drinking water, surface and ground water and waste-water for the presence of *E.coli* (2). MUG permits rapid detection of *E. coli* when medium is observed for fluorescence using UV Light (3, 4). MUG also detects anaerogenic strains which may not be detected in conventional procedure (3). MUG is hydrolyzed by the enzyme β-glucuronidase possessed by *E.coli* to yield a fluorescent end product 4-Methylumbelliferone. Tryptone provides essential nutrients. Lactose is the fermentable carbohydrate. Sodium chloride maintains osmotic equilibrium. The medium has a strong buffering system to control the pH in the presence of fermentative action. The bile salts inhibit gram-positive bacteria especially *Bacillus* species and faecal Streptococci. Mostly beta-glucuronidase activity occurs within 4 hours but some weak beta- glucuronidase-positive strains require overnight incubation (2). The fermentation of lactose by lactose fermentors leads to acidification of the medium, resulting in drop of pH. Adjustment of pH of cultures by sodium hydroxide enhanced fluorescence as observed by Maddocks and Greenman (5). Similarly Freir and Hartman (6) noticed that exposure of tubes to ammonia fumes enhanced fluorescence.

Nct i g"pw o dgt"qh"Proteus vulgaris"kh"rtgugpv." o c {"uwr rrtguu" i cu" rtqfwekqp"qh"E.coli." j q y xgt"hnwtguegpeg" rgt o kvu" fgvgvkqp" qh"E.coli"kp" rwtg"qt" o k z g f"ewnwvtgu" y kv jkp"6"vq"46"j qwtu0

Inoculate the test water sample into PA Broth (M1186) and Lauryl Sulphate Broth (M080). After an incubation at 35-37°C for 18-24 hours, all presumptive tubes showing growth, gas or acidity is further tested using MUG EC Broth (M1042). After an incubation at 35-37°C for 4-24 hours, the presence of a bright blue fluorescence is considered as a positive response for *E. coli* .

Quality Control

Appearance

Cream to yellow homogeneous free flowing powder

Colour and Clarity of prepared medium

Yellow coloured clear solution without any precipitate

Reaction

Please refer disclaimer Overleaf.

Reaction of 3.7% w/v aqueous solution at 25°C. pH : 6.9±0.2

pH

6.70-7.10

Cultural Response

M1042: Cultural characteristics observed after an incubation at 35 - 37°C for 4 - 24 hours.

Organism	Inoculum (CFU)	Growth	Fluorescence (under uv) (at 366 nm)
<i>Escherichia coli</i> ATCC 25922	50-100	luxuriant	positive, throughout the tube
<i>Enterobacter aerogenes</i> ATCC 13048	50-100	luxuriant	negative
<i>Staphylococcus aureus</i> ATCC 25923	>=10 ³	inhibited	
<i>Salmonella Typhi</i> ATCC 6539	50-100	good	negative
<i>Shigella flexneri</i> ATCC 12022	50-100	good	Negative

Storage and Shelf Life

Store below 30°C in tightly closed container and prepared medium below 2-8°C. Use before expiry period on the label.

Reference

- Hajna A. A. and Perry C. A., 1943, Am. J. Public Health, 33:550.
- Feng P. C. S. and Hartman P. A. S., 1982, Appl. Environ. Microbiol., 43:132.
- Robinson B. J., 1984, Appl. Environ. Microbiol., 48:285.
- Greenberg A. E., Trussell R. R. and Clesceri L. S., (Eds.), 1988, Standard Methods for the Examination of Water and Wastewater, 20th Ed., APHA, Washington, D.C.
- Maddocks J. L. and Greenan M. J. (1975) J. Clin. Pathol. 28. 686-687.
- Freir T. A. and Hartman P. A. (1987) Appl. Env. Microbiol. 53. 1246-1250.

Revision : 02 / 2015



Disclaimer :

User must ensure suitability of the product(s) in their application prior to use. Products conform solely to the information contained in this and other related HiMedia™ publications. The information contained in this publication is based on our research and development work and is to the best of our knowledge true and accurate. HiMedia™ Laboratories Pvt Ltd reserves the right to make changes to specifications and information related to the products at any time. Products are not intended for human or animal or therapeutic use but for laboratory, diagnostic, research or further manufacturing use only, unless otherwise specified. Statements contained herein should not be considered as a warranty of any kind, expressed or implied, and no liability is accepted for infringement of any patents.