



HiCrome™ Lactobacillus Selective Agar Base

M2065

Intended Use:

Recommended for isolation and differentiation between various species of *Lactobacillus* from a mixed culture by chromogenic method.

Composition**

Ingredients	Gms / Litre
Peptone	10.000
HM Extract #	1.000
M-Protein powder ##	5.000
D-Mannitol	10.000
Sodium chloride	10.000
Chromogenic mixture	3.200
Phenol red	0.025
Agar	15.000
Final pH (at 25°C)	7.1±0.2

**Formula adjusted, standardized to suit performance parameters

Equivalent to Meat Extract

Equivalent to Milk Protein

Directions

Suspend 54.22 grams in 1000 ml purified / distilled water. Heat to boiling to dissolve the medium completely. Sterilize by autoclaving at 15 lbs pressure (121°C) for 15 minutes. Cool to 45-50°C. Aseptically add rehydrated contents of 1 vial of Ciprofloxacin Supplement (FD345). Mix well and pour into sterile Petri plates.

Principle And Interpretation

Lactobacillus is a genus of Gram-positive, facultative anaerobic or microaerophilic, rod-shaped, non-spore-forming bacteria. They are a major part of the lactic acid bacteria group. As more LABs have been developed and sold in mixed forms as probiotics, it is necessary to develop a method for counting each LAB in a mixture (2).

The medium contains peptone and HM extract, which provide nitrogenous and carbonaceous compounds, long chain amino acids, vitamins and other essential nutrients. Mannitol serves as the fermentable carbohydrate, fermentation of which can be detected by phenol red. M-protein aids in detecting casein hydrolysis activity. The chromogenic mixture present in the medium is cleaved by the enzyme beta-glucosidase resulting in greenish blue to blue coloured colonies. For selective isolation of *Lactobacillus*, Ciprofloxacin supplement (FD345) is added which inhibits the accompanying bacteria.

Type of specimen

Clinical samples; Dairy samples: milk and milk products.

Specimen Collection and Handling

For clinical samples, follow appropriate techniques for sample collection and processing as per guidelines (3,4).

For dairy samples, follow appropriate techniques for sample collection and processing as per guidelines (1, 5).

After use, contaminated materials must be sterilized by autoclaving before discarding.

Warning and Precautions

In Vitro diagnostic Use. Read the label before opening the container. Wear protective gloves/protective clothing/eye protection/face protection. Follow good microbiological lab practices while handling specimens and culture. Standard precautions as per established guidelines should be followed while handling clinical specimens. Safety guidelines may be referred in individual safety data sheets.

Limitations

1. Some species may show poor growth due to nutritional variations.
2. Slight colour variation may be observed depending upon the utilization of the substrate by the organism.

Performance and Evaluation

Performance of the medium is expected when used as per the direction on the label within the expiry period when stored at recommended temperature.

Quality Control

Appearance

Light yellow to pink homogeneous free flowing powder

Gelling

Firm, comparable with 1.5% Agar gel

Colour and Clarity of prepared medium

Red coloured, clear to slightly opalescent gel forms in Petri plates

Reaction

Reaction of 5.42% w/v aqueous solution at 25°C. pH : 7.1±0.2

pH

6.90-7.30

Cultural Response

Cultural characteristics observed after an incubation at 25-30°C for 24-48 hours with addition of Ciprofloxacin Supplement (FD345)(with 5% CO₂).

Organism	Inoculum (CFU)	Growth	Recovery	Colour of Colony
<i>Lactobacillus acidophilus</i> ATCC 4356 (00098*)	50-100	good-luxuriant	≥50%	Pale pink - pink
<i>Lactobacillus casei</i> ATCC 9595	50-100	good-luxuriant	≥50%	Light green
<i>Lactobacillus fermentum</i> ATCC 9338	50-100	good-luxuriant	≥50%	Yellow
<i>Lactobacillus plantarum</i> ATCC 8014	50-100	good-luxuriant	≥50%	Light green-green colonies w/ hazy background
<i>Lactococcus lactis subsp. lactis</i> ATCC 19435 (00016*)	50-100	good-luxuriant	≥50%	Light green-green colonies w/ hazy background
<i>Bacillus subtilis subsp. spizizenii</i> ATCC 6633 (00003*)	≥10 ⁴	inhibited	0%	
<i>Staphylococcus aureus subsp. aureus</i> ATCC 6538 (00032*)	≥10 ⁴	inhibited	0%	
<i>Bacillus cereus</i> ATCC 10876	≥10 ⁴	inhibition	0%	

Key : (*) Corresponding WDCM numbers.

Storage and Shelf Life

Store between 15-25°C in a tightly closed container and the prepared medium at 2-8°C. Use before expiry date on the label. On opening, product should be properly stored dry, after tightly capping the bottle in order to prevent lump formation due to the hygroscopic nature of the product. Improper storage of the product may lead to lump formation. Store in dry ventilated area protected from extremes of temperature and sources of ignition Seal the container tightly after use. Product performance is best if used within stated expiry period.

Disposal

User must ensure safe disposal by autoclaving and/or incineration of used or unusable preparations of this product. Follow established laboratory procedures in disposing of infectious materials and material that comes into contact with sample must be decontaminated and disposed of in accordance with current laboratory techniques (3,4).

Reference

1. American Public Health Association, Standard Methods for the Examination of Dairy Products, 1978, 14th Ed., Washington D.C.
2. De Man, J.C., Rogosa, M. and Sharpe, E.M. (1960) A medium for the cultivation of lactobacilli. J Appl Bacteriol 23, 30–35.
3. Isenberg, H.D. Clinical Microbiology Procedures Handbook 2nd Edition.
4. Jorgensen, J.H., Pfaller, M.A., Carroll, K.C., Funke, G., Landry, M.L., Richter, S.S and Warnock., D.W. (2015) Manual of Clinical Microbiology, 11th Edition. Vol. 1.
5. Wehr H. M. and Frank J. H., 2004, Standard Methods for the Microbiological Examination of Dairy Products, 17th Ed., APHA Inc., Washington, D.C.


Revision : 00 / 2020



In vitro diagnostic medical device



CE Marking



Storage temperature



Do not use if package is damaged



HiMedia Laboratories Pvt. Limited,
23 Vadhani Industrial Estate,
LBS Marg, Mumbai-86, MS, India



CE Partner 4U, Esdoornlaan 13, 3951
DB Maarn The Netherlands,
www.cepartner4u.eu

Disclaimer :

User must ensure suitability of the product(s) in their application prior to use. Products conform solely to the information contained in this and other related HiMedia™ publications. The information contained in this publication is based on our research and development work and is to the best of our knowledge true and accurate. HiMedia™ Laboratories Pvt Ltd reserves the right to make changes to specifications and information related to the products at any time. Products are not intended for human or animal or therapeutic use but for laboratory, diagnostic, research or further manufacturing use only, unless otherwise specified. Statements contained herein should not be considered as a warranty of any kind, expressed or implied, and no liability is accepted for infringement of any patents.