

Limitations

As organisms differ in their nutritional requirements, some fastidious organisms may be inhibited or may show poor growth.

2. Further biochemical and serological tests must be carried out for complete identification.

Performance and Evaluation

Performance of the medium is expected when used as per the direction on the label within the expiry period when stored at recommended temperature.

Quality Control

Appearance

Cream to yellow homogeneous free flowing powder

Gelling

Firm, comparable with 1.5% Agar gel.

Colour and Clarity of prepared medium

Basal medium : Light amber coloured clear to slightly opalescent gel. After addition of 5% v/v sterile defibrinated blood : Cherry red coloured, opaque gel forms in Petri plates.

Reaction

Reaction of 5.2% w/v aqueous solution at 25°C. pH : 7.4±0.2.

pH

7.20-7.60

Cultural Response

Cultural characteristics observed after an incubation at 35-37°C for 18-24 hours (If desired add 5% v/v sterile defibrinated blood).

Organism	Inoculum (CFU)	Growth	Recovery	Growth w/ blood	Recovery w/ blood
<i>Candida albicans</i> ATCC 26790	50-100	luxuriant	≥70%	luxuriant	≥70%
<i>Staphylococcus aureus</i> subsp. <i>aureus</i> ATCC 25923 (00034*)	50-100	luxuriant	≥70%	luxuriant	≥70%
<i>Streptococcus pneumoniae</i> ATCC 6303	50-100	luxuriant	≥70%	luxuriant	≥70%
<i>Shigella flexneri</i> ATCC 12022 (00126*)	50-100	luxuriant	≥70%	luxuriant	≥70%
<i>Escherichia coli</i> ATCC 25922 (00013*)	50-100	luxuriant	≥70%	luxuriant	≥70%

Key : *Corresponding WDCM numbers.

Storage and Shelf Life

Store between 10- 30°C in a tightly closed container and the prepared medium at 2 - 8°C. Use before expiry date on the label. On opening, product should be properly stored dry, after tightly capping the bottle in order to prevent lump formation due to the hygroscopic nature of the product. Improper storage of the product may lead to lump formation. Store in dry ventilated area protected from extremes of temperature and sources of ignition Seal the container tightly after use. Use before expiry date on the label. Product performance is best if used within stated expiry period.

Disposal

User must ensure safe disposal by autoclaving and/or incineration of used or unusable preparations of this product. Follow established laboratory procedures in disposing of infectious materials and material that comes into contact with clinical sample must be decontaminated and disposed of in accordance with current laboratory techniques (7,8).

Reference

- Roseburg T. et al, 1944, J. Infect. Dis., 74:131.
- Conant N. F., 1950, Diagnostic Procedures and Reagents, 3rd Ed., APHA Inc.
- MacFaddin J. F., 1985, Media for Isolation-Cultivation-Identification-Maintenance of Medical Bacteria, Vol. I, Williams and Wilkins, Baltimore.
- Creitz and Puckett, 1954, Am. J. Clin. Pathol., 24:1318.
- Murray P. R., Baron J. H., Pfaller M. A., Jorgensen J. H. and Tenover F. C., (Eds.), 2003, Manual of Clinical Microbiology, 8th Ed., American Society for Microbiology, Washington, D.C.

6. Ajello L., Georg L., Kaplan W. and Kaufman L., 1963, CDC Laboratory Manual for Medical Mycology, PHS Publication No. 994, U.S. Govt. Office, Washington, D.C.
7. Isenberg, H.D. Clinical Microbiology Procedures Handbook 2nd Edition.
8. Jorgensen, J.H., Pfaller, M.A., Carroll, K.C., Funke, G., Landry, M.L., Richter, S.S and Warnock., D.W. (2015) Manual of Clinical Microbiology, 11th Edition. Vol. 1.


Revision : 04 / 2022



In vitro diagnostic medical device



CE Marking



Storage temperature



Do not use if package is damaged



HiMedia Laboratories Pvt. Limited,
Plot No.C-40, Road No.21Y,
MIDC, Wagle Industrial Area,
Thane (W)
-400604,MS,India



CE Partner 4U ,Esdoornlaan 13,
3951 DB Maarn The Netherlands,
www.cepartner4u.eu

Disclaimer :

User must ensure suitability of the product(s) in their application prior to use. Products conform solely to the information contained in this and other related HiMedia™ publications. The information contained in this publication is based on our research and development work and is to the best of our knowledge true and accurate. HiMedia™ Laboratories Pvt Ltd reserves the right to make changes to specifications and information related to the products at any time. Products are not intended for human or animal or therapeutic use but for laboratory, diagnostic, research or further manufacturing use only, unless otherwise specified. Statements contained herein should not be considered as a warranty of any kind, expressed or implied, and no liability is accepted for infringement of any patents.