



Violet Red Bile Glucose Agar w/o Lactose

M581

Intended Use:

Violet Red Bile Glucose Agar w/o Lactose is used for detection and enumeration of *Enterobacteriaceae* in raw food and clinical samples.

Composition**

Ingredients	Gms / Litre
Peptone	7.000
Yeast extract	3.000
Sodium chloride	5.000
Bile salts mixture	1.500
Glucose	10.000
Neutral red	0.030
Crystal violet	0.002
Agar	12.000
Final pH (at 25°C)	7.4±0.2

**Formula adjusted, standardized to suit performance parameters

Directions

Suspend 38.53 grams in 1000 ml purified / distilled water. Heat to boiling to dissolve the medium completely. DO NOT AUTOCLAVE. Cool to 45-50°C. Mix well and pour into sterile Petri plates.

Principle And Interpretation

Violet Red Bile Agar, a modification of MacConkey original formulation (7) is used for the enumeration of coli-aerogenes bacterial group.

Violet Red Bile Glucose Agar w/o Lactose, a modification of VRBA (M049), was designed for the enumeration of *Enterobacteriaceae* (9). It employs the selective inhibitory components crystals violet and bile salts and the indicator system glucose and neutral red. Sought bacteria will dissimilate glucose and produce purple zones around the colonies (3). ISO committee has also recommended this medium (4). Selectivity of VRBGA can be increased by incubation under anaerobic conditions and/ or at elevated temperature, i.e. equal to or above 42°C (10-12). Peptone and yeast extract serve as sources of carbon, nitrogen, vitamins and other essential growth nutrients. Glucose is the fermentable carbohydrate, utilization of which leads to the production of acids. Neutral red indicator detects the acidity so formed. Crystal violet and bile salts mixture help to inhibit the accompanying gram-positive and unrelated flora. Sodium chloride maintains the osmotic equilibrium. Further biochemical tests are necessary for positive identification (8).

Type of specimen

Clinical samples; Food and dairy samples; Water samples

Specimen Collection and Handling

For clinical samples follow appropriate techniques for handling specimens as per established guidelines (5,6).

For food and dairy samples, follow appropriate techniques for sample collection and processing as per guidelines (1,13,14).

For water samples, follow appropriate techniques for sample collection, processing as per guidelines and local standards(2).

After use, contaminated materials must be sterilized by autoclaving before discarding.

Warning and Precautions

In Vitro diagnostic use. Read the label before opening the container. Wear protective gloves/protective clothing/eye protection/face protection. Follow good microbiological lab practices while handling specimens and culture. Standard precautions as per established guidelines should be followed while handling clinical specimens. Safety guidelines may be referred in individual safety data sheets.

Limitations

1. Over incubation may result in reverting of reaction.
2. Further biochemical tests must be carried out for confirmation.

Performance and Evaluation

Performance of the medium is expected when used as per the direction on the label within the expiry period when stored at recommended temperature.

Quality Control

Appearance

Light yellow to pinkish beige homogeneous free flowing powder

Gelling

Firm, comparable with 1.2% Agar gel.

Colour and Clarity of prepared medium

Reddish purple coloured clear to slightly opalescent gel forms in Petri plates.

Reaction

Reaction of 3.85% w/v of aqueous solution at 25°C. pH : 7.4±0.2

pH

7.20-7.60

Cultural Response

Cultural characteristics was observed after an incubation at 35-37°C for 18-24 hours. Recovery rate is considered as 100% for bacteria growth on Soyabean Casein Digest Agar.

Cultural Response

Organism	Inoculum (CFU)	Growth	Recovery	Colour of colony
Cultural Response				
<i>Escherichia coli</i> ATCC 25922 (00013*)	50 -100	good-luxuriant	≥50 %	pink-red
<i>Escherichia coli</i> ATCC 8739 (00012*)	50 -100	luxuriant	≥50 %	pink-red with bile precipitate
<i>Escherichia coli</i> NCTC 9002	50 -100	luxuriant	≥50 %	pink-red with bile precipitate
<i>Pseudomonas aeruginosa</i> ATCC 9027 (00026*)	50 -100	good-luxuriant	≥50 %	pink to red
<i>Salmonella Enteritidis</i> ATCC 13076 (00030*)	50 -100	good-luxuriant	≥50 %	light pink
# <i>Klebsiella aerogenes</i> ATCC 13048 (00175*)	50 -100	good-luxuriant	≥50 %	pink-red
<i>Staphylococcus aureus</i> subsp.aureus ATCC 25923 (00034*)	≥10 ⁴	inhibited	0%	
<i>Staphylococcus aureus</i> subsp.aureus ATCC 6538 (00032*)	≥10 ⁴	inhibited	0%	

Key : *Corresponding WDCM numbers.

- Formerly known as *Enterobacter aerogenes*

Storage and Shelf Life

Store between 10-30°C in a tightly closed container and the prepared medium at 20-30°C. Use before expiry date on the label. On opening, product should be properly stored dry, after tightly capping the bottle in order to prevent lump formation due to the hygroscopic nature of the product. Improper storage of the product may lead to lump formation. Store in dry ventilated area protected from extremes of temperature and sources of ignition Seal the container tightly after use. Use before expiry date on the label. Product performance is best if used within stated expiry period.

Disposal

User must ensure safe disposal by autoclaving and/or incineration of used or unusable preparations of this product. Follow established laboratory procedures in disposing of infectious materials and material that comes into contact with clinical sample must be decontaminated and disposed of in accordance with current laboratory techniques (5,6).

Reference

- 1.American Public Health Association, Standard Methods for the Examination of Dairy Products, 1978, 14th Ed., Washington D.C.
- 2.Baird R.B., Eaton A.D., and Rice E.W., (Eds.), 2015, Standard Methods for the Examination of Water and Wastewater, 23rd ed., APHA, Washington, D.C.
- 3.Corry J. E. L., Curtis G. D. W. and Baird R. M., (Ed.), 1995, Culture Media for Food Microbiology, Vol. 34, Progress in Industrial Microbiology, Elsevier, Amsterdam.
- 4.International Organization for Standardization (ISO), 1993, Draft ISO/DIS 7402.
- 5.Isenberg, H.D. Clinical Microbiology Procedures Handbook. 2nd Edition.
- 6.Jorgensen,J.H., Pfaller , M.A., Carroll, K.C., Funke, G., Landry, M.L., Richter, S.S and Warnock., D.W. (2015) Manual of Clinical Microbiology, 11th Edition. Vol. 1.
- 7.MacConkey A., 1905, J. Hyg., 5, 333-379.
- 8.MacFaddin J. F., 1985, Media for Isolation-Cultivation-Identification-Maintenance of Medical Bacteria, Vol. 1, Williams and Wilkins, Baltimore
- 9.Mossel D. A. A., Eclderink I., Koopmans M. and Van Rossem F., 1978, Lab. practice, 27 No. 12: 1049.
- 10.Mossel D. A. A. and Vega C. L., 1973, Hlth. Lab. Sci., 11:303
- 11.Mossel D. A. A., Eclderink I., Koopmans M. and Van Rossem F., 1979, Food Protect., 42 : 470
- 12.Mossel D. A. A. et al, 1986, J. Appl. Bacteriol., 60:289.
- 13.Salfinger Y., and Tortorello M.L. Fifth (Ed.), 2015, Compendium of Methods for the Microbiological Examination of Foods, 5th Ed., American Public Health Association, Washington, D.C.
- 14.Wehr H. M. and Frank J. H., 2004, Standard Methods for the Microbiological Examination of Dairy Products, 17th Ed., APHA Inc., Washington, D.C.


Revision : 03/ 2018



In vitro diagnostic medical device



CE Marking



Storage temperature



Do not use if package is damaged



HiMedia Laboratories Pvt. Limited,
23 Vadhani Industrial Estate,
LBS Marg,Mumbai-86,MS,India



CE Partner 4U ,Esdoornlaan 13, 3951
DB Maarn The Netherlands,
www.cepartner4u.eu

Disclaimer :

User must ensure suitability of the product(s) in their application prior to use. Products conform solely to the information contained in this and other related HiMedia™ publications. The information contained in this publication is based on our research and development work and is to the best of our knowledge true and accurate. HiMedia™ Laboratories Pvt Ltd reserves the right to make changes to specifications and information related to the products at any time. Products are not intended for human or animal or therapeutic use but for laboratory, diagnostic, research or further manufacturing use only, unless otherwise specified. Statements contained herein should not be considered as a warranty of any kind, expressed or implied, and no liability is accepted for infringement of any patents.