



## Littman Bile Broth Base

M663

### Intended Use:

Recommended for selective enrichment and cultivation of pathogenic fungi.

### Composition\*\*

Ingredients	Gms / Litre
Peptone	10.000
Dextrose (Glucose)	10.000
Bile #	15.000
Crystal violet	0.010
Final pH ( at 25°C)	7.0±0.2

\*\*Formula adjusted, standardized to suit performance parameters

#Equivalent to Oxgall

### Directions

Suspend 35.01 grams in 1000 ml purified/distilled water. Heat, if necessary, to dissolve the medium completely. Sterilize by autoclaving at 15 lbs pressure (121°C) for 15 minutes. Cool to 45-50°C and aseptically add sterile Streptomycin to a final concentration of 30 mcg/ml of medium. Mix well and dispense as desired.

### Rtkpekrng"Cpf"Kpvgtrtgvcvkqp

Nkw o cp"Qz i cm"Dtqv j" Dcug" y cu" hqt o wncv g f" d{ "Nkv o cp"\*4.5+0" Nkv o cp"Qz i cm"Dtqv j" Dcug" ku" wug f" hqt" ugngev kxg" gptkej o gpv" qh" rcvjqigpke" umkp" hwpik" \*fgt o cvqrj {vgu+ " cpf" ucrtqrj {vke" hwpik" htqo" xctkquw" enpkecn" urgek o gpu0" Kv" rtqxfgu" ghhev kxg" gptkej o gpv" gxgp" y jgp" vjg" vguw" uc o rngu" ctg" jgcxkn{ "eqpvc o kpevgf" ykvj" dcevgtkcn" hntc0" Nkv o cp"Qz i cm" o gfkc" ctg" cnuq" wug f" hqt" vjg" gpw o gtcvkqp" qh" hwpicn" rqrwncvkqp" qh" ckt. "uqkn. "hqqfuvwhhu" cpf" qv jgt" o cvgtkcn" qh" ucpkvct{ "uki pkhkecpeg"\*6+0

Crystal violet and Streptomycin has inhibitory effect on most of the bacteria. Bile restricts spreading of fungal colonies. The neutral pH favours the growth of many pathogenic fungi.

For inoculation, skin or nail scraping or infected hair is directly placed on the surface of Littman Bile Agar Base (M373) while sputum, faeces etc. are spread over the surface with sterile swab or the specimen are first enriched in broth and then cultured on agar medium. Whenever *Nocardia asteroides*, *Streptomyces* or any Streptomycin sensitive microorganisms are to be cultured, use the medium without Streptomycin (4).

### Type of specimen

Clinical samples - skin or nail scraping or infected hair etc

### Specimen Collection and Handling:

For clinical samples follow appropriate techniques for handling specimens as per established guidelines (1,2).

After use, contaminated materials must be sterilized by autoclaving before discarding.

### Warning and Precautions

In Vitro diagnostic Use. Read the label before opening the container. Wear protective gloves/protective clothing/eye protection/ face protection. Follow good microbiological lab practices while handling specimens and culture. Standard precautions as per established guidelines should be followed while handling clinical specimens. Safety guidelines may be referred in individual safety data sheets.

### Limitations :

1. The incubation should be carried out for upto 8 days.

### Performance and Evaluation

Performance of the medium is expected when used as per the direction on the label within the expiry period when stored at recommended temperature.

## Quality Control

### Appearance

Cream to yellow homogeneous free flowing powder

### Colour and Clarity of prepared medium

Blue coloured clear solution in tubes

### Reaction

Reaction of 3.5% w/v aqueous solution at 25°C. pH : 7.0±0.2

### pH

6.80-7.20

### Cultural Response

Cultural characteristics observed with added sterile Streptomycin to a final concentration of 30mcg/ml of medium, after an incubation at 25-30°C for 48-72 hours .

Organism	Growth (Plain Growth with medium)	Streptomycin
<i>Aspergillus flavus</i> ATCC 22547	luxuriant	good-luxuriant
<i>Candida albicans</i> ATCC 10231 (00054*)	good - luxuriant	good - luxuriant
<i>Escherichia coli</i> ATCC 25922 (00013*)	good - luxuriant	inhibited
<i>Microsporium audouinii</i> ATCC 9079	luxuriant	good-luxuriant
<i>Saccharomyces cerevisiae</i> ATCC 9763 (00058*)	good - luxuriant	good - luxuriant
<i>Saccharomyces uvarum</i> ATCC 28098	good - luxuriant	good - luxuriant
<i>Trichophyton mentagrophytes</i> ATCC 9533	good	good

Key : \* - Corresponding WDCM numbers.

## Storage and Shelf Life

Store between 10-30°C in a tightly closed container and the prepared medium at 20-30°C. Use before expiry date on the label. On opening, product should be properly stored dry, after tightly capping the bottle in order to prevent lump formation due to the hygroscopic nature of the product. Improper storage of the product may lead to lump formation. Store in dry ventilated area protected from extremes of temperature and sources of ignition. Seal the container tightly after use.

Product performance is best if used within stated expiry period.

## Disposal

User must ensure safe disposal by autoclaving and/or incineration of used or unusable preparations of this product. Follow established laboratory procedures in disposing of infectious materials and material that comes into contact with clinical sample must be decontaminated and disposed of in accordance with current laboratory techniques (1,2).

## Reference

1. Isenberg, H.D. Clinical Microbiology Procedures Handbook 2<sup>nd</sup> Edition.
2. Jorgensen, J.H., Pfaller, M.A., Carroll, K.C., Funke, G., Landry, M.L., Richter, S.S and Warnock., D.W. (2015) Manual of Clinical Microbiology, 11th Edition. Vol. 1.
2. Littman M. L., 1947, Science, 106:109.
3. Littman M. L., 1948, Am. J. Clin. pathol., 18:409.
4. MacFaddin J. F., 1985, Media for Isolation-Cultivation-Identification-Maintenance of Medical Bacteria, Vol.1, Williams and Wilkins, Baltimore.


Revision : 02 / 2020



In vitro diagnostic medical device



CE Marking



Storage temperature



Do not use if package is damaged



HiMedia Laboratories Pvt. Limited,  
23 Vadhani Industrial Estate,  
LBS Marg, Mumbai-86, MS, India



CE Partner 4U, Esdoornlaan 13, 3951  
DB Maarn The Netherlands,  
[www.cepartner4u.eu](http://www.cepartner4u.eu)

**Disclaimer :**

User must ensure suitability of the product(s) in their application prior to use. Products conform solely to the information contained in this and other related HiMedia™ publications. The information contained in this publication is based on our research and development work and is to the best of our knowledge true and accurate. HiMedia™ Laboratories Pvt Ltd reserves the right to make changes to specifications and information related to the products at any time. Products are not intended for human or animal or therapeutic use but for laboratory, diagnostic, research or further manufacturing use only, unless otherwise specified. Statements contained herein should not be considered as a warranty of any kind, expressed or implied, and no liability is accepted for infringement of any patents.