



Technical Data

HiCrome™ UTI HiCynth™ Agar, Modified

MCD1418

Intended use: n t n

HiCrome™ UTI HiCynth™ Agar, Modified is chromogenic differential medium for identification, differentiation and confirmation of enteric bacteria from specimens such as urine which may contain large number of *Proteus* species as well as potentially pathogenic gram-positive organisms.

Composition**

Ingredients	Gms / Litre
HiCynth peptone No.2#	18.000
HiCynth peptone No.5#	10.000
Chromogenic mixture	12.440
Agar	15.000
Final pH (at 25°C)	7.2±0.2

**Formula adjusted, standardized to suit performance parameters

Directions

Suspend 55.44 grams in 1000 ml purified / distilled water. Heat to boiling to dissolve the medium completely. Sterilize by autoclaving at 15 lbs pressure (121°C) for 15 minutes. Cool to 45-50°C. Mix well and pour into sterile Petri plates.

Principle And Interpretation

HiCrome™ UTI Agar, Modified is formulated on the basis of work carried out by Pezzlo (7), Wilkie et al (11), Friedman et al (2), Murray et al (6), Soriano and Ponte (9 and Merlino et al (5). These media is the modification of HiCrome™ UTI Agar (M1353), which can be used in place of MacConkey Agar for isolation and confirmation of various microorganisms. It facilitates and expedites the identification of some gram-negative bacteria and some gram-positive bacteria on the basis of different contrasted colony colours produced by reactions of genus or species specific enzymes with two chromogenic substrates. HiCrome™ UTI HiCynth™ Agar, Modified is prepared by completely replacing animal or vegetable based peptones with chemically defined peptones to avoid BSE/TSE and GMO risks associated with vegetable peptones.

Enzymes produced by *Enterococcus* species, *Escherichia coli* and coliforms cleave the chromogenic substrates incorporated in the medium. Presence of rich source of phenylalanine and tryptophan from peptone and tryptone provides an indication of tryptophan deaminase activity, revealed with TDA Reagent (R036) indicating the presence of *Proteus* species, *Morganella* species and *Providencia* species, which appear brown. One chromogenic substrate is cleaved by β-glucosidase possessed by Enterococci resulting in formation of blue colonies. *E.coli* produce purple-magenta colonies due to the enzyme β-D-galactosidase which cleaves the other chromogenic substrate. Further confirmation of *E.coli* can be done by performing indole test using DMACA Reagent (R035). Also, some strains of *Enterobacter cloacae* lacking β-glucosidase show pink-colonies indistinguishable from *E.coli*. The DMACA reagent for indole test (should be performed on filter paper) distinguishes between *E.coli* and *Enterobacter*, and also between *Proteus mirabilis* and other species. Coliforms produce purple coloured colonies due to cleavage of both the chromogenic substrates HiCynth™ peptone No.2 and HiCynth™ peptone No.5 provides nitrogenous, carbonaceous compounds and other essential growth nutrients.

Type of specimen

Clinical samples : urine, faeces , Food samples , Water samples.

Specimen Collection and Handling

For clinical samples follow appropriate techniques for handling specimens as per established guidelines (4,6).

For food and dairy samples, follow appropriate techniques for sample collection and processing as per guidelines (8,10).

For water samples, follow appropriate techniques for sample collection, processing as per guidelines and local standards.(1)

After use, contaminated materials must be sterilized by autoclaving before discarding.

Warning and Precautions

In Vitro diagnostic use. Read the label before opening the container. Wear protective gloves/protective clothing/eye protection/face protection. Follow good microbiological lab practices while handling specimens and culture. Standard precautions as per established guidelines should be followed while handling clinical specimens. Safety guidelines may be referred in individual safety data sheets.

Limitations

1. Since it is an enzyme-substrate based reaction, the intensity of colour may vary with isolates.

Performance and Evaluation

Performance of the medium is expected when used as per the direction on the label within the expiry period when stored at recommended temperature.

Quality Control

Appearance

Cream to yellow homogeneous free flowing powder

Gelling

Firm, comparable with 1.5% Agar gel

Colour and Clarity of prepared medium

Light amber coloured, clear to slightly opalescent gel forms in Petri plates

Reaction

Reaction of 5.54% w/v aqueous solution at 25°C. pH : 7.2±0.2

pH

7.00-7.40

Cultural Response

Cultural characteristics observed after an incubation at 35-37°C for 24 hours.

Organism	Inoculum (CFU)	Growth	Recovery	Colour of Colony	TDA (add 1-2 drops of TDA reagent)	DMACA (transfer colony on filter paper dipped in DMACA Reagent)
<i>Escherichia coli</i> ATCC 25922 (00013*)	50-100	luxuriant	≥70%	Purple to magenta	negative reaction	positive reaction, formation of blue purple colour around growth
<i>Enterococcus faecalis</i> ATCC 29212 (00087*)	50-100	luxuriant	≥70%	blue-green (small)	negative reaction	negative reaction
<i>Klebsiella pneumoniae</i> ATCC 13883 (00097*)	50-100	luxuriant	≥70%	blue to purple, mucoid	negative reaction	negative reaction
<i>Proteus mirabilis</i> ATCC 12453	50-100	luxuriant	≥70%	light brown	positive reaction, development of brown colouration	negative reaction
<i>Pseudomonas aeruginosa</i> ATCC 27853 (00025*)	50-100	luxuriant	≥70%	colourless (greenish pigment may be observed)	negative reaction	negative reaction
<i>Staphylococcus aureus</i> subsp. <i>aureus</i> ATCC 25923 (00034*)	50-100	luxuriant	≥70%	golden yellow	negative reaction	negative reaction

Key : *Corresponding WDCM numbers.

Storage and Shelf Life

Store between 15-25°C in a tightly closed container and the prepared medium at 2-8°C. Use before expiry date on the label. On opening, product should be properly stored dry, after tightly capping the bottle in order to prevent lump formation due to the hygroscopic nature of the product. Improper storage of the product may lead to lump formation. Store in dry ventilated area protected from extremes of temperature and sources of ignition Seal the container tightly after use. Product performance is best if used within stated expiry period.

Disposal

User must ensure safe disposal by autoclaving and/or incineration of used or unusable preparations of this product. Follow established laboratory procedures in disposing of infectious materials and material that comes into contact with clinical sample must be decontaminated and disposed of in accordance with current laboratory techniques (3,4).

Reference

1. Baird R.B., Eaton A.D., and Rice E.W., (Eds.), 2015, Standard Methods for the Examination of Water and Wastewater, 23rd ed., APHA, Washington, D.C.
2. Friedman M.P. et al. (1991), Journal of Clinical Microbiology, 29:2385-2389.
3. Isenberg, H.D. Clinical Microbiology Procedures Handbook. 2nd Edition.
4. Jorgensen, J.H., Pfaller, M.A., Carroll, K.C., Funke, G., Landry, M.L., Richter, S.S and Warnock., D.W. (2015) Manual of Clinical Microbiology, 11th Edition. Vol. 1.
5. Merlino et al. (1995), Abstr. Austr. Microbiol., 16(4):17-3.
6. Murray P., Tenover P. and Tenover F.C., (1992), Journal of Clinical Microbiology, 30:1600-1601.
7. Pezzlo M, (1998), Clinical Microbiology Reviews, 1:268-280
8. Salfinger Y., and Tortorello M.L., 2015, Compendium of Methods for the Microbiological Examination of Foods, 5th Ed., American Public Health Association, Washington, D.C.
9. Soriano F. and Ponte C., (1992), Journal of Clinical Microbiology, 30:3033-3034.
10. Wehr H. M. and Frank J. H., 2004, Standard Methods for the Microbiological Examination of Dairy Products, 17th Ed., APHA Inc., Washington, D.C.
11. Wilkie M.E., Almond M.K. and Marsh F.P., (1992), British Medical Journal, 305:1137-1141.


Revision : 02 / 2020



In vitro diagnostic medical device



CE Marking



Storage temperature



Do not use if package is damaged



HiMedia Laboratories Pvt. Limited,
23 Vadhani Industrial Estate,
LBS Marg, Mumbai-86, MS, India



CE Partner 4U ,Esdoornlaan 13, 3951
DB Maarn The Netherlands,
www.cepartner4u.eu

Disclaimer :

User must ensure suitability of the product(s) in their application prior to use. Products conform solely to the information contained in this and other related HiMedia™ publications. The information contained in this publication is based on our research and development work and is to the best of our knowledge true and accurate. HiMedia™ Laboratories Pvt Ltd reserves the right to make changes to specifications and information related to the products at any time. Products are not intended for human or animal or therapeutic use but for laboratory, diagnostic, research or further manufacturing use only, unless otherwise specified. Statements contained herein should not be considered as a warranty of any kind, expressed or implied, and no liability is accepted for infringement of any patents.