



## Leifson HiVeg™ Agar

MV1380

### Intended use

Recommended for isolation of *Salmonella* and *Shigella* species from clinical and non-clinical samples.

### Composition\*\*

Ingredients	Gms / Litre
HiVeg™ extract No. 1	5.000
HiVeg™ peptone No. 1	5.000
Lactose	10.000
Sodium thiosulphate	5.400
Sodium citrate	6.000
Ferric citrate	1.000
Synthetic detergent No. III	3.000
Neutral red	0.020
Agar	12.000
Final pH ( at 25°C)	7.5±0.2

\*\*Formula adjusted, standardized to suit performance parameters

### Directions

Suspend 47.42 grams in 1000 ml purified / distilled water. Heat to boiling to dissolve the medium completely. DO NOT AUTOCLAVE OR OVERHEAT. Excessive heating is detrimental. Cool to 45-50°C. Mix well and pour into sterile Petri plates.

### Principle And Interpretation

*Salmonella* and *Shigella* are gram-negative, facultatively anaerobic, non-sporulating, non-motile rods in the family *Enterobacteriaceae*. They are widely distributed in animals affecting mainly the stomach and the intestines. Leifson Agar is recommended for isolation of *Salmonella* and *Shigella* species (6). Leifson HiVeg™ Agar is same as Leifson Agar except that the animal based peptones are completely replaced with vegetable peptones to avoid the BSE/TSE risks associated with animal peptones.

HiVeg™ extract No. 1 and HiVeg™ peptone No. 1 provides nitrogenous and carbonaceous compounds, long chain amino acids and other essential growth nutrients. Synthetic detergent No. III inhibit all gram-positive bacteria. Lactose is added to the medium to allow differentiation of lactose fermenting bacteria such as *Escherichia coli* from non-lactose fermenting species such as *Salmonella* and *Shigella* species. Lactose fermenting strains grow as red to pink colonies because of absorption of neutral red indicator. Sodium thiosulphate and ferric citrate forms the H<sub>2</sub>S indicator system. Non-fermenting species grow as colorless colonies with black centers due to production of H<sub>2</sub>S against *Shigella* which does not produce H<sub>2</sub>S (7).

### Type of specimen

Clinical samples - Blood; Food and dairy samples; Water samples.

### Specimen Collection and Handling

For clinical samples follow appropriate techniques for handling specimens as per established guidelines (4,5).

For food and dairy samples, follow appropriate techniques for sample collection and processing as per guidelines (1,3,8).

For water samples, follow appropriate techniques for sample collection, processing as per guidelines and local standards.(3)

After use, contaminated materials must be sterilized by autoclaving before discarding.

### Warning and Precautions :

In Vitro diagnostic Use. Read the label before opening the container. Wear protective gloves/protective clothing/eye protection/ face protection. Follow good microbiological lab practices while handling specimens and culture. Standard precautions as per established guidelines should be followed while handling clinical specimens. Safety guidelines may be referred in individual safety data sheets.

### Limitations :

1. This medium is general purpose medium and may not support the growth of fastidious organisms.

## Performance and Evaluation

Performance of the medium is expected when used as per the direction on the label within the expiry period when stored at recommended temperature.

## Quality Control

### Appearance

Light yellow to pink homogeneous free flowing powder

### Gelling

Firm, comparable with 1.2% Agar gel.

### Colour and Clarity of prepared medium

Reddish orange coloured clear to slightly opalescent gel forms in Petri plates

### Reaction

Reaction of 4.74% w/v aqueous solution at 25°C. pH : 7.5±0.2

### pH

7.30-7.70

### Cultural Response

MV1380: Cultural characteristics observed after an incubation at 35-37°C for 18-24 hours.

Organism Cultural Response	Inoculum (CFU)	Growth	Recovery	Colour of Colony	H <sub>2</sub> S
<i>Escherichia coli</i> ATCC 25922 (00013*)	50-100	poor	10-20%	pink w/bile precipitate	negative reaction
<i>Enterococcus faecalis</i> ATCC 29212 (00087*)	>=10 <sup>4</sup>	inhibited	0 %		
<i>Salmonella</i> Enteritidis ATCC 13076 (00030*)	50-100	good-luxuriant	>=50%	colourless	positive reaction, black centred colonies
<i>Shigella flexneri</i> ATCC 12022 (00126*)	50-100	good	40-50%	colourless	negative reaction
<i>Salmonella</i> Typhimurium ATCC 14028 (00031*)	50-100	luxuriant	>=50%	colourless	positive reaction, black centred colonies

## Storage and Shelf Life

Store between 10-30°C in a tightly closed container and the prepared medium at 2-8°C. Use before expiry date on the label. On opening, product should be properly stored dry, after tightly capping the bottle in order to prevent lump formation due to the hygroscopic nature of the product. Improper storage of the product may lead to lump formation. Store in dry ventilated area protected from extremes of temperature and sources of ignition Seal the container tightly after use. Use before expiry date on the label.

Product performance is best if used within stated expiry period.

## Disposal

User must ensure safe disposal by autoclaving and/or incineration of used or unusable preparations of this product. Follow established laboratory procedures in disposing of infectious materials and material that comes into contact with clinical sample must be decontaminated and disposed of in accordance with current laboratory techniques (4,5).

## Reference

1. American Public Health Association, Standard Methods for the Examination of Dairy Products, 1978, 14th Ed., Washington D.C.
2. Downes F. P. and Ito K., (Ed.), 2001, Compendium of Methods for the Microbiological Examination of Foods, 4th Ed., American Public Health Association, Washington, D.C.
3. Greenberg A. E., Clesceri L. S. and Eaton A. D., (Eds.), 2005, Standard Methods for the Examination of Water and Wastewater, 21st ed., APHA, Washington, D.C.
4. Isenberg, H.D. Clinical Microbiology Procedures Handbook. 2nd Edition.
5. Jorgensen, J.H., Pfaller, M.A., Carroll, K.C., Funke, G., Landry, M.L., Richter, S.S and Warnock., D.W. (2015) Manual of Clinical Microbiology, 11th Edition. Vol. 1.

6. Leifson, E., 1935, J. Pathol. Bacteriol., 40-581.
7. Macfaddin J. 1985, Media for Isolation-Cultivation-Identification-Maintenance of Medical Bacteria, Vol.1. Williams and Wilkins, Baltimore.
8. Wehr H. M. and Frank J. H., 2004, Standard Methods for the Microbiological Examination of Dairy Products, 17th Ed., APHA Inc., Washington, D.C.


Revision : 01 / 2019



In vitro diagnostic medical device



CE Marking



Storage temperature



Do not use if package is damaged



HiMedia Laboratories Pvt. Limited,  
23 Vadhani Industrial Estate,  
LBS Marg, Mumbai-86, MS, India



CE Partner 4U ,Esdoornlaan 13, 3951  
DB Maarn The Netherlands,  
[www.cepartner4u.eu](http://www.cepartner4u.eu)

#### Disclaimer :

User must ensure suitability of the product(s) in their application prior to use. Products conform solely to the information contained in this and other related HiMedia™ publications. The information contained in this publication is based on our research and development work and is to the best of our knowledge true and accurate. HiMedia™ Laboratories Pvt Ltd reserves the right to make changes to specifications and information related to the products at any time. Products are not intended for human or animal or therapeutic use but for laboratory, diagnostic, research or further manufacturing use only, unless otherwise specified. Statements contained herein should not be considered as a warranty of any kind, expressed or implied, and no liability is accepted for infringement of any patents.