



Kovac's Indole Reagent

R008

Intended use

For detection of presence of indole produced by microorganisms due to tryptophan deamination.

Composition**

Ingredients	-
p-dimethylamino benzaldehyde	5.000
Amyl alcohol	75.000
Hydrochloric acid, concentrated	25.000

**Formula adjusted, standardized to suit performance parameters

Directions

Add 0.2 - 0.3 ml of Kovac's reagent to 5 ml of a 24 - 48 hours old culture of the organism under investigation. Formation of a red coloured ring indicates positive indole test.

Principle And Interpretation

Peptone Water is particularly suitable as a substrate in the study of indole production. Peptone used in Peptone Water, is rich in tryptophan content (1). Other peptones which contain tryptophan can be used to study indole production. Tryptone Water is also recommended by APHA (2) for detection of indole production by coliforms, which is a key feature in differentiation of bacteria. It is used as part of the IMViC procedures. Most strains of *E. coli*, *P. vulgaris*, *P. rettgeri*, *M. morgani* and *Providencia* species break down the amino acid tryptophan with the release of indole. The presence of indole can be detected by the addition of Ehrlich's or Kovac's reagent (p-dimethylaminobenzaldehyde).

Kovacs reagent is a biochemical reagent consisting of isoamyl alcohol, para-dimethylaminobenzaldehyde (DMAB), and concentrated hydrochloric acid. It is used for the diagnostic test, to determine the ability of the organism to split tryptophan into indole and alpha-aminopropionic acid by hydrolytic activity of bacteria that express tryptophanase enzyme (3). The indole produced is indicated by formation of a red coloured ring, soluble in ether, chloroform and alcohol. This was invented by the Hungarian-Swiss Chemist, Ervin Kovacs. Indole production is used as, a tests designed to distinguish among members of the family Enterobacteria.

Type of specimen

Clinical samples ; Water samples

Specimen Collection and Handling

For clinical samples follow appropriate techniques for handling specimens as per established guidelines (4,5).

For water samples, follow appropriate techniques for sample collection, processing as per guidelines and local standards(2).

After use, contaminated materials must be sterilized by autoclaving before discarding.

Warning and Precautions :

In Vitro diagnostic Use only. Read the label before opening the container. Wear protective gloves/protective clothing/eye protection/ face protection. Follow good microbiological lab practices while handling specimens and culture. Standard precautions as per established guidelines should be followed while handling clinical specimens. Safety guidelines may be referred in individual safety data sheets

Limitations :

1. Growth media must contain an adequate amount of tryptophan. Do not use Mueller- Hinton Agar for test, because tryptophan is destroyed during the acid hydrolysis of casein.
2. Do not used media that contain dyes (e.g., EMB, MAC).
3. Do not use medium with a nitrate disc/strip to perform the indole test, as nitrate can interfere with indole test by including false positive results.

Performance and Evaluation

Performance of the product is expected when used as per the direction on the label within the expiry period when stored at recommended temperature.

Quality Control

Appearance

Greenish yellow coloured solution

Solubility

Immiscible with water

Clarity

Clear with no insoluble particles.

Cultural Response

Characteristic reactions observed when Kovac's Indole Reagent is added to growth in Tryptone Broth (M463)

Organism	Indole production
<i>Enterobacter aerogenes</i> ATCC 13048	negative reaction ,no red ring
<i>Escherichia coli</i> ATCC 25922	positive reaction, red ring at the interface of the medium

Storage and Shelf Life

Store between 10-30°C in tightly closed container and away from bright light. Use before expiry date on label. On opening, product should be properly stored in dry ventilated area protected from extremes of temperature and sources of ignition. Seal the container tightly after use.

Disposal

User must ensure safe disposal by autoclaving and/or incineration of used or unusable preparations of this product. Follow established laboratory procedures in disposing of infectious materials and material that comes into contact with clinical sample must be decontaminated and disposed of in accordance with current laboratory techniques (4,5).

Reference


1. MacFaddin J., 1980, Biochemical Tests for Identification of Medical Bacteria, 2nd ed., Williams and Wilkins, Baltimore.
2. Greenberg A. E., Clesceri L. S. and Eaton A. D., (Eds.), 2005, Standard Methods for the Examination of Water and Wastewater, 21st ed., APHA, Washington, D.C.
3. MacFaddin J. F., 2000, Biochemical Tests for Identification of Medical Bacteria, 3rd Ed., Williams and Wilkins, Baltimore.
4. Isenberg, H.D. Clinical Microbiology Procedures Handbook. 2nd Edition.
5. Jorgensen, J.H., Pfaller, M.A., Carroll, K.C., Funke, G., Landry, M.L., Richter, S.S and Warnock, D.W. (2015) Manual of Clinical Microbiology, 11th Edition. Vol. 1.



In vitro diagnostic medical device



CE Marking



Storage temperature



Do not use if package is damaged



HiMedia Laboratories Pvt. Limited,
B /4-6 , MIDC, Dindori, Nashik MH

www.himedialabs.com



CE Partner 4U ,Esdoornlaan 13, 3951
DB Maarn The Netherlands,
www.cepartner4u.eu



Disclaimer :

User must ensure suitability of the product(s) in their application prior to use. Products conform solely to the information contained in this and other related HiMedia™ publications. The information contained in this publication is based on our research and development work and is to the best of our knowledge true and accurate. HiMedia™ Laboratories Pvt Ltd reserves the right to make changes to specifications and information related to the products at any time. Products are not intended for human or animal or therapeutic use but for laboratory, diagnostic, research or further manufacturing use only, unless otherwise specified. Statements contained herein should not be considered as a warranty of any kind, expressed or implied, and no liability is accepted for infringement of any patents.